

IFU Size : 161 H x 104 W

CARVENE GUTTA PERCHA SOLVENT



Product Description : Carvene Gutta Percha Solvent is used for removing gutta percha at the time of root canal re-treatment. Also useful for removing all types of root canal filling materials including Zinc Oxide – Eugenol Cement or as a cleaner for endodontic instruments. Carvene is non-hazardous unlike Chloroform/Trichloroethylene which are considered as carcinogenic substances. Also it is less volatile as compared to chloroform / Trichloroethylene. Carvene Gutta Percha Solvent is based on natural substances and therefore it does not cause irritation to oral mucosa and peri-apical tissue.

Composition : d-Limonene which is a natural substance extracted from orange oil.

Intended Purpose / Intended Use : It is intended for removal of gutta-percha / root canal filling materials

Indications for Use :

1. Solvent for removal of Gutta-percha
2. Solvent for removal of zinc oxide eugenol based root canal sealers
3. Solvent for cleaning endodontic instruments

Instructions for Use : Using the dropper, drop 1- 2 drops of liquid in the root canal or in close proximity to the material to be removed. Remove gutta-percha / root canal filling materials with an appropriate root canal instrument. Re-apply if the gutta-percha / root canal filling is dry or crumbly. If required, use another dose. Repeat until the desired effect is reached. The dispensed material shall not be placed back in container and the dose once applied and used shall not be reused.

Warnings : Avoid prolonged or repeated contact with skin (allergic contact dermatitis), and eyes. Avoid prolonged inhalation. Do not take internally. Contact with the skin may cause irritation, burns of hypersensitivity. If contact with skin occurs immediately wipe off thoroughly with cotton and alcohol and then wash well with soap and water. If skin rash and sensitization or other allergic reaction occurs discontinue use and seek medical attention. If the product comes into contact with the eyes, rinse thoroughly with fresh water and consult a physician. It should not be used with patients who have a history of severe allergic reaction to any of the components.

Precautions : Carvene Gutta Percha Solvent is volatile and flammable. Keep away from heat source. Keep bottle tightly capped after every use.

Contraindications :

1. The product should not be used with patients with a history of allergic reaction to any of its components.
2. The product should not be used beyond the root apex.

Storage Conditions : Store at temperature between 10°C(50°F) to 24°C(75°F). Close the cap carefully after use. Keep away from moisture. Shelf life not affected after opening if stored in a cool dry place.

Shelf life : 3 years from the date of manufacturing.

Presentation :

Ref : 40030

1 x 10ml Glass Bottle, 1 x Dispensing Dropper.

Disposal conditions : To be disposed off based on the prevailing rules of the country in which it is used.

Prevest DenPro Limited
Unit II, EPIP, Bari Brahmana,
Jammu-181133, India.
www.prevestdenpro.com
Customer Care No. : + 91 8899074151
Email : customercare@prevestdenpro.com
Made in India

EC REP
Obelis s.a
Boulevard General Wehls 53
1030 Brussels, Belgium
Tel : +32 2 739 59 54, Fax : +32 2 739 60 93
E-mail : mail@obelis.net

CE
0123

Revision : 09
Date : 06-01-2023

UDI Symbol for "UNIQUE DEVICE IDENTIFIER"

LOT Symbol for "BATCH CODE" USE-BY DATE Symbol for "USE-BY DATE" NON STERILE Symbol for "NON STERILE" KEEP DRY Symbol for "KEEP DRY"

CAUTION Symbol for "CAUTION" CONSULT INSTRUCTION FOR USE Symbol for "CONSULT INSTRUCTION FOR USE" REF Symbol for "CATALOGUE NUMBER"

EC REP Symbol for "AUTHORIZED REPRESENTATIVE IN THE EUROPEAN COMMUNITY/EUROPEAN UNION"

KEEP OUT OF THE REACH OF CHILDREN Symbol for "KEEP OUT OF THE REACH OF CHILDREN" MANUFACTURER Symbol for "MANUFACTURER" DO NOT USE IF PACKAGE IS DAMAGE Symbol for "DO NOT USE IF PACKAGE IS DAMAGE"

TEMPERATURE LIMIT Symbol for "TEMPERATURE LIMIT" KEEP AWAY FROM SUNLIGHT Symbol for "KEEP AWAY FROM SUNLIGHT" MD Symbol for "MEDICAL DEVICE"

The material has been developed solely for professional dental use. Application should be carried out strictly according to the instruction for use. Liability cannot be accepted for damages resulting from failure to observe the instructions of the stipulated area of applications. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in this instruction sheet. Description and data constitute no warranty of attributes and are not binding. For feedback, complaint and reporting of unexpected events contact customer care.