CITROGEN

(20% Citric Acid Liquid)

Dentine conditioner and chelating agent

PREVESTDenPro®
THE FUTURE OF DENTISTRY

Product description : Citrogen is a 20 % Citric Acid liquid used as a Chelating agent for removal of smear layer and conditioning of dentine surfaces.

Composition : Citric acid, DM water, Preservatives & Stablizers.

Intended use: It is used as a Chelating agent for removal of smear layer and conditioning of dentine surfaces.

Indications for use :

· Conditioning of prepared root canal systems
· Removal of smear layer
· Lubrication during root canal preparation
· Conditioning of dentine surfaces

Instructions for use:

For Root canal procedures: Irrigate the root canal with the product using a leur lock syringe coupled with a side vented irrigation needle. Do not extrude the liquid beyond the apex. Adequate care should be taken to restrict the material within the canal. After using the product, rinse the canal with physiological saline. Citrogen is packaged in a multiple dose container (Bottle). Opening and closing of the container does not affect the shelf life of the device. The dispensed material shall not be placed back in the container and the dose once applied or used shall not be reused.

Conditioning of dentine surfaces: Apply the liquid on the dentinal surfaces to the primed or conditioned . Brush it lightly with a Microbrush and leave it in place for a period of 3-5 minutes. Rinse the preparation with physiological saline.

Warnings :

1. The material may cause sensitization in susceptible people. Do not use Citrogen in people allergic to any components of this product.
2. Avoid eye contact to prevent irritation. In case of contact with eyes, rinse immediately with plenty of water and seek medical attention. Avoid skin contact to prevent possible allergic response. Material may cause irritation and therefore contact with skin and mucous membrane should be avoided .In case of contact, immediately remove material with cotton and wash thoroughly with water and soap. In case of skin sensitization or rash, discontinue the use and seek medical attention.
3. Do not swallow or take internally. If accidental swallowing occurs, drink lot of water. If nausea or illness develops, seek medical attention.

Precautions :

1. Always use a rubber dam while working with this product.
2. Do not push the solution beyond the apex.
3. Use gloves, masks and protective eyewear while handling the solution to avoid exposure.

Contraindications : Do not use the product for patients with hypersensitivity to the product's components. Do not leave the product in the canal or the tooth cavity between visits. This may result in excessive softening of dentin. Restrict the use of the product to a maximum of 5 minutes, when used as a conditioning agent

Storage conditions : This product should be used at room temperature. Contact with water should be strictly avoided. Inadequate storage conditions could shorten shelf life and product performance. We recommend to store at temperatures between 10°C(50°F) and 24°C(75°F) and keep syringe cap tightly closed. .

Shelf Life : 2 years from the date of manufacturing.

Packaging :

Ref : 40047
1 x 50ml Bottle, 1 x Dropper
Ref : 40048
1 x 100ml Bottle, 1 x Dropper

Disposal conditions: To be disposed off based on the prevailing rules of the country in which it is used.

Prevest DenPro Limited
Unit II, Export Promotion Industrial Park,
(EPIP), Bari Brahmana, Jammu-181133,
India. www.prevestdenpro.com
Customer Care No. : + 91 8899074151
Email : customercare@prevestdenpro.com
Made in India

EC REP
Obelis s.a
Boulevard General Wahis 53
1030 Brussels, Belgium
Tel: +(32) 2 732 59 54, Fax. +(32) 2 732 60 03
E-mail : mail@obelis.net

CE
0123

Revision :	03
Date :	28-07-2025

UDI Symbol for "UNIQUE DEVICE IDENTIFIER"

LOT Symbol for "BATCH CODE" Symbol for "USE BY" NON STERILE Symbol for "NON STERILE" Symbol for "KEEP DRY"

Symbol for "CAUTION" Symbol for "CONSULT INSTRUCTION FOR USE" REF Symbol for "CATALOGUE NUMBER"

EC REP SYMBOL FOR EUROPEAN COMMUNITY AUTHORIZED REPRESENTATIVE Symbol for "DO NOT USE IF PACKAGE IS DAMAGE"

Symbol for "KEEP OUT OF THE REACH OF CHILDREN" Symbol for "MANUFACTURER NAME & ADDRESS"

Symbol for "TEMPERATURE LIMITATION" Symbol for "KEEP AWAY FROM SUNLIGHT" MD Symbol for "MEDICAL DEVICE"

The material has been developed solely for professional dental use. Application should be carried out strictly according to the instruction for use. Liability cannot be accepted for damages resulting from failure to observe the instructions of the stipulated area of applications. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in this instruction sheet. Description and data constitute no warranty of attributes and are not binding. For feedback, complaint and reporting of unexpected events contact customer care.